



Republic of the Philippines  
**LAGUNA LAKE DEVELOPMENT AUTHORITY**

**BID BULLETIN NO. 1**

This supplemental/bid bulletin is issued to inform all participating bidders of the following amendments to the issued bidding documents for the **supply and delivery of various Laboratory Equipment (Rebidding) with Contract No. LAB-2021-05:**

Item No.	Equipment	Technical Specifications	
		Original	Amendment/Additional Specifications
3	Dispensette	<ul style="list-style-type: none"> <li>- Discharge tube with or without recirculation valve</li> <li>- Calibration Certificate</li> </ul>	<ul style="list-style-type: none"> <li>- Preferably with recirculation valve.</li> <li>- Certificate of Calibration to be provided during delivery of equipment, not during the opening of bids</li> </ul>
4	Magnetic Stirrer	<ul style="list-style-type: none"> <li>- Dimensions: (WxDxH) approx. 185x185x65mm</li> <li>- Stirring quantity: 2L max</li> </ul>	<ul style="list-style-type: none"> <li>- Dimensions: (WxDxH) approx. 185x185x65mm (minimum)</li> <li>- Stirring quantity: 1L to 2L</li> </ul>
5	DO Meter	<ul style="list-style-type: none"> <li>- Temperature Range: -30.0 to 130 °C</li> </ul>	<ul style="list-style-type: none"> <li>-Temperature Range: 0 - 50 °C</li> </ul>
6	pH Meter	<ul style="list-style-type: none"> <li>- electrode strand</li> <li>- Temperature range: -5 to 105 deg C</li> <li>- Display: LED</li> </ul>	<ul style="list-style-type: none"> <li>- With electrode stand</li> <li>- Benchtop</li> <li>- Temperature range: 0 - 100 °C</li> <li>- Display: LED or LCD</li> </ul>

Please be guided accordingly.

  
**ENGR. EMITERIO C. HERNANDEZ, EnP**  
 BAC Chairperson

**ISO 9001:2015 Certified Quality Management System**  
 National Ecology Center, East Avenue, Diliman, 1100, Quezon City  
 Tel. Nos. (02) 8376-4072, (02) 8376-4039, (02) 8376-4044, (02) 8376-5430, (02) 8376-5433  
 Website: <http://www.llda.gov.ph> | E-mail: [info@llda.gov.ph](mailto:info@llda.gov.ph)

